

Volume	Test volume	Increments	Inaccuracy ¹	Imprecision ²	Cat. No.
0.1-2.5 µL	0.25 µL	0.002 µL	12.00 %	6.00 %	8-100-30-9
	1.25 µL		2.50 %	1.50 %	
	2.5 µL		2.50 %	1.50 %	
0.5-10 µL	1 µL	0.02 µL	2.50%	1.50%	8-101-30-9
	5 µL		1.50%	0.80%	
	10 µL		1.00%	0.40%	
2-20µL	2 µL	0.02 µL	3.00%	1.50%	8-102-30-9
	10 µL		1.20%	0.60%	
	20 µL		0.90%	0.30%	
5-50 µL	5 µL	0.10 µL	2.00%	2.00%	8-103-30-9
	25 µL		0.80%	0.40%	
	50 µL		0.60%	0.30%	
10-100 µL	10 µL	0.10 µL	3.00%	1.00%	8-104-30-9
	50 µL		1.00%	0.30%	
	100 µL		0.80%	0.20%	
20-200 µL	20 µL	0.20 µL	2.50%	0.70%	8-105-30-9
	100 µL		0.70%	0.30%	
	200 µL		0.60%	0.20%	
100-1000 µL	100 µL	1.00 µL	3.00%	0.60%	8-106-30-9
	500 µL		1.00%	0.20%	
	1000 µL		0.60%	0.20%	
0.5-5 mL	0.5 mL	10 µL	2.40%	0.60%	8-107-30-9
	2.5 mL		1.20%	0.25%	
	5 mL		0.60%	0.20%	
1-10 mL	1 mL	20 µL	3.00%	0.60%	8-108-30-9
	5 mL		0.80%	0.20%	
	10 mL		0.60%	0.15%	

¹ Systematic error
² Random error